





UNLESS OTHERWISE SPECIFIED MANUFACTURE PER ALLEN INSTITUTE WORKMANSHIP STANDARD

MINIMUM CONTENT DRAWING. MANUFACTURE TO 3D MODEL WITHIN +/- 0.25mm UNLESS OTHERWISE SPECIFIED. REFER TO 3D MODEL FOR UNSPECIFIED DIMENSIONS.

REVISIONS			
REV.	DESCRIPTION	DATE	APRVD





Ø 3.18[0.125] THRU ALL



Ø 6.76[0.266] THRU ALL

Ø 5.11 THRU ALL  
1/4-20 UNC THRU ALL

MATERIAL 6061 Alloy	UNLESS OTHERWISE SPECIFIED:  DIMENSIONS ARE IN MILLIMETERS TOLERANCES:  DECIMAL    ± .1            ANGULAR: MACH ± 2°  BREAK ALL EDGES .5 MAX INTERPRET PER ASME Y14.5-2009 MAX ROUGHNESS 1.6 micrometers	PROPRIETARY AND CONFIDENTIAL CONTACT ALLEN INSTITUTE FOR RULES REGARDING HANDLING		 ALLEN INSTITUTE for BRAIN SCIENCE <i>Fueling Discovery</i>	615 Westlake Ave N Seattle, WA 98109    USA Phone 206-548-7000    www.alleninstitute.org		
FINISH		DRAWN <i>davidsu</i>	DATE 4/5/2018		TITLE: Surgical Tooling, Headframe Placement		
		APPROVED	DATE				
TREATMENT		COMMENTS:		SIZE <b>B</b>	DWG. NO. <b>0111-200-01</b>	REV <b>X01</b>	
COLOR		THIRD ANGLE PROJECTION 		DO NOT SCALE DRAWING		SCALE: 1:1	WEIGHT (g): 0.050.05